

Return & Repair Form:

Customer Information:

Company Name: Prompton Lakes B: _____ S: _____
 Date Received: 10-28-09 Date Given to Service: _____
 Carrier: FedEx UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-(1110)-6510-6810-7510 Serial #: AR XA-0033
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FL

Part Number	Description	Qty	Total Cost
<u>6808453</u>	<u>FL</u>	<u>1</u>	<u>N/C WARR</u>
<u>6809458</u>	<u>DETEK TO</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP CCL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP LA 10</u>	<u>LABO</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes:
FL ARWJ-0057 Replaced with ARAE-1182
CAL W/OE; Ops Check

Service Technician: JS

Date: 10/30/09