

RETURN AND REPAIR FORM



Customer Information

B: _____ S: _____

Company Name: Bloomfield PD NJ

Date Received: 7-28-11 Date given to service: 7-28-11

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR XA-0034

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information

Test#: _____

Reason for Return: _____

Part Number	Description	Qty	Total Cost
<u>mp Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C W</u>
<u>mp Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

Repair Notes: Unit was evaluated and tested, after ABA Test Failure unit was opened and evaluated, Tech saw that the hose from the cuvette to the pressure sensor was not connected, hose was connected and unit then passed ABA tests. Fuel Cell AR BE-1182

Cal w/ QC & ops check

Service Technician: [Signature] Date: 7-28-11