

Return & Repair Form:

Customer Information:

Company Name: BLOOMINGDALE P.D. B: _____ S: _____
 Date Received: 1-5-09 Date Given to Service: 1-5-09
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 110 - 6510 - 6810 Serial #: AR XA - 0035
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FC + IR

| Part Number | Description | Qty | Total Cost |
|------------------|----------------------------------|-------------|-----------------|
| <u>6808455</u> | <u>FC</u> | <u>1</u> | <u>N/C WARR</u> |
| <u>MP MISC</u> | <u>SERVICE NON STOCK CUVERTE</u> | <u>1</u> | <u>N/C WARR</u> |
| <u>MP CAL 71</u> | <u>CAL</u> | <u>1</u> | <u>N/C WARR</u> |
| | | | |
| | | | |
| | | | |
| | | | |
| <u>MP LABOR</u> | | <u>1 HR</u> | <u>N/C WARR</u> |

Repair Notes:

REPLACED FC IN BLOOMINGDALE DETECTOR SIDE WINDOW ON CUVERTE /
REPLACED CUVERTE W/ SERVICE CUVERTE /

CAL W/ QC & OPS CHECKS

Service Technician: [Signature]

Date: 01-07-09