

Return & Repair Form:

Customer Information:

Company Name: Bloomfield PD B: _____ S: _____
 Date Received: 1/6/10 Date Given to Service: 1/6/10
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo
 Serial #: ARXA - 0035
 Printer Ser #: AR -
 Whole Instrument _____ Top 1/2 _____
 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:			
<u>IR Detektor / FC</u>			
Part Number	Description	Qty	Total Cost
<u>6809458</u>	<u>IR Detektor</u>	<u>1</u>	<u>n/k w/wh</u>
<u>6809455</u>	<u>FC</u>	<u>1</u>	<u>n/k w/wh</u>
<u>MP CAL 41</u>	<u>CAL</u>	<u>1</u>	<u>n/k w/wh</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>n/k w/wh</u>

Repair Notes:

INSTALL New IR-Detektor / FC

CAL w/DC / ops check

Service Technician: AS

Date: 01/08/2010