

RETURN AND REPAIR FORM



Customer Information

B: _____ S: _____

Company Name: North Caldwell P.D., North Caldwell, NJ

Date Received: 07/22/13 Date given to service: 07/22/13

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810 Serial Number: AR XA-0038
 7110 8610 Printer Serial#: AR
 7510 DT5000 Sim Serial#: DD
 7410 Upper-half 9510 Probe Serial#: DD
 7410 Whole

Warranty Expires: December 2013

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information

Test#: _____

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
12003	Hose Barb, 1/8" Outlet	1	NC/W
MP Labor	Labor	5	NC/W

Repair Notes: Replaced defective hose barb, 1/8" NV PM Boy 7110 CS Outlet.
Performed multiple tests to re-duplicate the issues for the orange button and could not duplicate it.

CAL W/QC AND OPS CHECK

Service Technician BC Date: 07/30/13