

# Return & Repair Form

## Customer Information:

Company Name: Bradley Beach P.D. B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 10-24-08 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR X 17 - 0039  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -

## ACCESSORIES:

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: E23 / FC

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>6809458</u>	<u>IR DETECTOR PCB</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP MISC</u>	<u>IR FILTER</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP LABOR</u>		<u>.5</u>	<u>N/C WARR</u>

## Repair Notes:

REPLACE FC / REPLACE IR DETECTOR / REPLACE IR FILTER - GLUE WAS ON FILTER - / CAL W/ QC & OPS CHECKS

Service Technician: \_\_\_\_\_

Date: 10.27.08