

## Return & Repair Form

### Customer Information:

CUSTOMER Sussex County Sheriff's Cust # 60162  
Date of Return: 11/20/06 Carrier: Fed Ex Method: priority

### Product Information:

Product : 7110 7410 6510 6810 Serial # ARXA-0040  
Description: A B PLUS SCREENER \_ Printer Serial # \_\_\_\_\_  
Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
OTHER : \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
Demo Unit  Returned to stock on: \_\_\_\_\_  
Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: ERROR 112

Part #	Description	Qty	Total
<u>IMPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Cleared error 112. Verified units accuracy with ops check. Full cal ops check FOC.

Technician: CRD

Date: 11/21/06