

Return & Repair Form

Customer Information:

CUSTOMER SUSSEX CO. Sheriff Cust #
Date of Return: 4-30-07 Carrier: FX Method: SAVEN

Product Information:

Product: 7110 7410 6510 6810 Serial # ARXA-0040
Description: A B PLUS SCREENER Printer Serial #
Whole Inst. Top Half Simulator Serial #
OTHER: Probe Serial #
Demo Unit Returned to stock on:

Accessories: (check all that apply):

- 110VAC Adapter Regulator Mag Card Rdr #
Printer Paper Printer Ribbon Casio #
Mouthpieces Carrying Case Dry Gas
Other (specify):

Warranty Expiration Date:

Repair Information:

Test #

Reason for Return: EC Problems

Table with 4 columns: Part #, Description, Qty, Total. Rows include 6808455 Fuel Cell, MPLABOR LABOR, and MPCAL71 7110 CAL.

Repair Notes: Replaced questionable FC. Full cal ops check + QC.

Technician: CRD

Date: 5/1/07