

RETURN AND REPAIR FORM

Customer Information

B:

S:

Company Name: Sussex County Sheriff, Newton, NJ

Date Received: 06/28/13 Date given to service: 06/28/13

Carrier: FedEx UPS USPS Shipping Method: GRD NDA-PRI 3DAY 2DAY NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR XA-0040

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

| Part Number | Description | Qty | Total Cost |
|-------------|---------------|-----|------------|
| MP Cal 71 | Calibration | 1 | NC/W |
| 6808455 | Fuel Cell | 1 | NC/W |
| 6808486 | Spacer Plates | 2 | NC/W |
| MP Labor | Labor | .5 | NC/W |
| | | | |
| | | | |
| | | | |
| | | | |

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = AREE-0277

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 07/25/13