

# Return & Repair Form:

## Customer Information:

Company Name: NORTH ARLINGTON P.D.B. S: \_\_\_\_\_  
 Date Received: 10.8.09 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR XA - 0042  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: <u>NO BEEP - - - AGAIN</u>			
Part Number	Description	Qty	Total Cost
<u>8315075</u>	<u>PCB</u>	<u>1</u>	<u>n/c w/whr</u>
<u>MPCAC 71</u>	<u>CAL</u>	<u>1</u>	<u>n/c w/whr</u>
<u>MPLABOR</u>		<u>1 HR</u>	<u>n/c w/whr</u>

## Repair Notes:

REPLACED MAIN PCB / EXCHANGE TOP COVER WITH DEMO UNIT (TOP INCLUDES NEW WIRE HARNESS, HORN, DISPLAY, PRINTER AND BUTTON PCB)

CAL W/ QC AND OPS CHECKS

Service Technician: [Signature]

Date: 10/14/2009