

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Woodland Park Police Station , Woodland Park NJ

Date Received: 03-05-12 Date given to service: 03-05-12

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810 Serial Number: AR XA-0043  
 7110  8610 Printer Serial#: AR  
 7510  DT5000 Sim Serial#: DD  
 7410 Upper-half  9510 Probe Serial#: DD  
 7410 Whole

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
8315075	PCB Board	1	N/C W
MP CAL 71	Calibration	1	N/C W
MP Labor	Labor	.5	N/C W

Repair Notes: Customer was not able to properly download there data from the unit, Installed a new (PCB) .

Cal W/QC AND OPS CHECK

Service Technician JS Date: 04-24-12