

# Return & Repair Form:

## Customer Information:

Company Name: ABSECON PD B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 1-29-09 Date Given to Service: 1-29-09  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR XA - 0045  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR - \_\_\_\_\_  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P - \_\_\_\_\_

**ACCESSORIES:**

110 V A/C Adapter       Regulator       Mag Card Rdr# \_\_\_\_\_  
 Printer Paper       Printer Ribbon       Casio # \_\_\_\_\_  
 Mouthpieces       Carrying Case       Dry Gas

Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: PROBLEM W/ LIN TEST

Part Number	Description	Qty	Total Cost
<u>MPCAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C W/ MKR</u>

## Repair Notes:

FC IS XN (DEC '06) /  
CLEAR MEMORY  
PERFORM STANDARD CAL W/ QC & OPS CHECKS PROCESS  
PERFORM LIN TEST = OK

Service Technician: [Signature]

Date: 01-30-09