

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Absecon City P.V., Absecon, NJ

Date Received: 2/3/12 Date given to service: 2/3/12

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR XA-0045
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
<u>6803455</u>	<u>Fuel Cell</u>	<u>1</u>	<u>N/C W</u>
<u>mp cal 71</u>	<u>Calibration</u>	<u>1</u>	<u>N/C W</u>
<u>mp Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

Repair Notes: Replaced Fuel cell # ACBD-2535 would not pass calibration test, new Fuel cell ACK-1075 JS ACK-1075.

cal w/QC & ops check

Service Technician JS Date: 2-9-12