

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Absecon City P.D., Absecon, NJ

Date Received: 06/12/14 Date given to service: 06/12/14

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR XA-0045  
 Printer Serial#: AR  
 Sim Serial#: DD  
 Probe Serial#: DD

Warranty Expires: May 2016

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W
12030	Cermetek Modem Chip	1	NC/W

Repair Notes: Replaced defective cermetek modem chip, C156, C167, fuel cell, and spacer plates.  
Customer reported that the modem would not connect and would be sporadic. Changing of the  
cermetek modem chip, C156, and C167 corrected this issue. Ran multiple tests with the modem and  
met with positive results.

Fuel Cell = ARFD-1843  
 CAL W/QC AND OPS CHECK

Service Technician BC Date: 06/17/14