

Return & Repair Form:

Customer Information:

Company Name: 87th SFS / JOINT BASE PDB: _____ S: _____
 Date Received: 2-24 Date Given to Service: 2-25
 Carrier: FedEx - UPS - DHL - (USPS) Method: GRD - (NDA) - 2DY - 3DY - Other

Product Information:

Product: 7410-~~7110~~-6510-6810-7510 Serial #: AR XA-0046
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -

ACCESSORIES:

110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas w/ Ranch Corp
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FL

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FL</u>	<u>1</u>	
<u>6808486</u>	<u>Plate</u>	<u>1</u>	
<u>mp Cal 71</u>	<u>Cal</u>	<u>1</u>	
<u>mp labor</u>	<u>Labor</u>	<u>.5</u>	

Repair Notes:

FL # ARWJ-0063 replaced with FL # ARBW-0841
We also had to wait for the sergeant to call back
to do a data download. Download on 3/4/2011
Cal w/ Qc, ops check

Service Technician: [Signature]

Date: 3/7/2011