

## RETURN AND REPAIR FORM

### Customer Information

B:

S:

Company Name: Sparta Twp. P.D., Sparta, NJ

Date Received: 06/21/13

Date given to service: 06/21/13

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR XA-0050

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: May 2015

Description:  A  B  Plus  Demo  Screener  Trade In

### Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

### Repair Information:

Test#:

| Part Number | Description | Qty | Total Cost |
|-------------|-------------|-----|------------|
| MP Cal 71   | Calibration | 1   | NC/W       |
| MP Labor    | Labor       | .5  | NC/W       |

Repair Notes: Resoldered a loose breath hose wire.

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 06/24/13