

Return & Repair Form:

Customer Information:

Company Name: Boro of Wildwood Crest B: _____ S: _____
 Date Received: 3-15-11 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-110-6510-6810-7510 Serial #: AR XA -0051
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FC

Part Number	Description	Qty	Total Cost
<u>6808955</u>	<u>FC</u>	<u>1</u>	<u>n/c w</u>
<u>mq cal 71</u>	<u>Cal</u>	<u>1</u>	<u>n/c w</u>
<u>6808986</u>	<u>Plate</u>	<u>1</u>	<u>n/c w</u>
<u>mq labor</u>	<u>Labor</u>	<u>15</u>	<u>n/c w</u>

Repair Notes:

FC# ARBM-3269

cal w/DC, ops check

Service Technician: LE

Date: 3/17/2011