

Return & Repair Form:

Customer Information:

Company Name: Stockton Police ^{collec} B: _____ S: _____
 Date Received: 2-9-10 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-~~9110~~-6510-6810-7510
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo
 Serial #: AR XA - 0052
 Printer Ser #: AR -
 Whole Instrument Top ½ Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:			
<u>FC</u>			
Part Number	Description	Qty	Total Cost
<u>6808453</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C WARR</u>
<u>6809500</u>	<u>CUVETTE</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP Labo</u>	<u>Labo</u>	<u>15</u>	<u>N/C WARR</u>

Repair Notes:

NEW FLT# ARAM-0766
Clean SPOTS off MIRROR of CUVETTE.
CHECK FRV 3.997 Bump Test PASSED.
Cal w/DC i OPS Check 2/16/10
JOE advised Tech to install new CUVETTE. 2/15/10
RAW 99 STD-Check on WARR. 2/15/10

Service Technician: JS

Date: 2/16/2010