

1
NJ

Dräger safety

Internal Use Only ✓

Return & Repair Form

Customer Information:

CUSTOMER OCPD Ocean City Police Cust # 62337
 Date of Return: 7/11/07 Carrier: FedEx Method: P1

Product Information:

Product # 7110 7410 6510 6810 Serial # ARXA-0059
 Description: A B PLUS SCREENER _ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER : _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): POWERCORD
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: FC PROB.

Part #	Description	Qty	Total
6808455	FUEL CELL	1	warrant
8100284	PRINT RIBBON	1	N/C
MPCAC 71	7110 Ctl	1	NC
MPLABOR		.5	N/C

Repair Notes: new fuel cell 8RXN 1583

Technician: VMB

Date: 07-12-07