

Return & Repair Form:

**Customer Information:**

Company Name: Occursity B: 60337 S: \_\_\_\_\_  
 Date Received: 6-29-10 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

**Product Information:**

Product: 7410-7110-6510-6810-7510 Serial #: AR XA-0057  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date YES

**Repair Information:**

Test #

Reason for Return: FC/IR Detector

Part Number	Description	Qty	Total Cost
<u>6805455</u>	<u>FC</u>	<u>1</u>	<u>N/C w</u>
<u>6809458</u>	<u>IR Detector</u>	<u>1</u>	<u>N/C w</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C w</u>
<u>MP Lab</u>	<u>Labo</u>	<u>.5</u>	<u>N/C w</u>

Repair Notes:

FC # RR117-2513

Installed new cap tube

CAL w/DC, ops check

Service Technician: \_\_\_\_\_

Date: 6/30/2010