

Return & Repair Form:

Customer Information:

Company Name: AVALON P.D. B: _____ S: _____
 Date Received: 12-18-08 Date Given to Service: 12-18-08
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY Other

Product Information:

Product: 7410 7110 - 6510 - 6810 Serial #: ARXA - 0059
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P - _____
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Reason for Return: IR PROBLEM Test # _____

Part Number	Description	Qty	Total Cost
<u>6809500</u>	<u>CUVETTE</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPC1C71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MV LABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes:
REPLACED CUVETTE / FC WAS INITIALED 9-11-08 (ZE-SERIES)
CAL w/ QC & OPS CHECKS

Service Technician: [Signature] Date: 12/19/08