

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Montclair P.D., Montclair, NJ

Date Received: 01/22/2018

Date given to service: 01/22/2018

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR XA-0063

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

## Test#:

| Part Number | Description | Qty | Total Cost |
|-------------|-------------|-----|------------|
| 4414161     | Calibration | 1   |            |
| 4414166     | Labor       | 1   |            |
| 6808455     | Fuel Cell   | 1   |            |
|             |             |     |            |
|             |             |     |            |
|             |             |     |            |
|             |             |     |            |

Repair Notes: Replaced defective fuel cell.

Complaint of modem issues, was unable to duplicate but changed C156 and C167 as a precaution.

Fuel Cell = ARKL-1273

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 03/12/2018