

# Dräger safety

## Return & Repair Form

### Customer Information:

CUSTOMER Passaic County Sheriff Dept - Cust # \_\_\_\_\_  
 Date of Return: 6/28/06 Carrier: Fed ex Method: Priority

### Product Information:

Product: 7110 7410 6510 6810 Serial # ARXA-0064  
 Description: A B PLUS SCREENER \_ Printer Serial # \_\_\_\_\_  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER: \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
 Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: IR readings randomly high

Part #	Description	Qty	Total

Repair Notes: Could not duplicate random high IR readings after 174 tests. Continued proper operation of 7110. ops check & Q.C.

Technician: CRD

Date: 6/30/06