

# Return & Repair Form

## Customer Information:

Company Name: Passaic County Sheriff Dept. B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 4-1-08 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410 7110 - 6510 - 6810 Serial #: AR XA - 0064 <sup>Priority</sup>  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR - \_\_\_\_\_  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD \_\_\_\_\_ P - \_\_\_\_\_  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: IR RESULTS

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP MISC</u>	<u>IR DETECT (PN 680945B) NOW STOCK</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCML 71</u>	<u>CML</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>1.0</u>	<u>N/C WARR</u>

### Repair Notes:

REPLACE FC + IR DETECTOR / CAL W/ QC'S OPS CHECKS

---



---



---



---



---

Service Technician: [Signature]

Date: 04-03-08