

# Dräger safety

## Return & Repair Form

### Customer Information:

CUSTOMER Hamilton Twp PD. Cust # 602703/602704  
 Date of Return: 2/15/17 Carrier: Fed Ex Method: Priority

### Product Information:

Product : 7110 7410 6510 6810 Serial # ARXA-0065  
 Description: A B PLUS SCREENER Printer Serial # \_\_\_\_\_  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER : \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
 Demo Unit  Returned to stock on: \_\_\_\_\_ **★ NOT PACKAGED WELL**  
 Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: repair ?? ask Joe was sent to EC Drops on his attention  
SOLD-CHANGE

| Part #         | Description      | Qty       | Total |
|----------------|------------------|-----------|-------|
| <u>W808455</u> | <u>Fuel Cell</u> | <u>1</u>  |       |
| <u>MPLABOR</u> | <u>LABOR</u>     | <u>.5</u> |       |
| <u>MPCAL71</u> | <u>7110 CAL</u>  | <u>1</u>  |       |
|                |                  |           |       |
|                |                  |           |       |

Repair Notes: Replaced questionable F/C. Full cal ops check + QC.

Technician: CRD

Date: 2/19/07