

RETURN AND REPAIR FORM

Customer Information

B:

S:

Company Name: Longport P.D., Longport, NJDate Received: 12/18/13Date given to service: 12/18/13Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STDProduct: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 WholeSerial Number: AR XA-0067Printer Serial#: ARSim Serial#: DDProbe Serial#: DDWarranty Expires: May 2016Description: A B Plus Demo Screener Trade In

Accessories

 110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
MP Labor	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.Fuel Cell = AREL-0392CAL W/QC AND OPS CHECKService Technician BCDate: 01/06/14