

# Return & Repair Form

## Customer Information:

Company Name: MARLBORO T/S B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 11-3-08 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - (NDA) - 2DY - 3DY - Other

## Product Information:

Product: 7410 - (7110) - 6510 - 6810 Serial #: AR XA - 0069  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR - \_\_\_\_\_  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P - \_\_\_\_\_

**ACCESSORIES:**

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: FC? - ERROR 32 NOTED

Part Number	Description	Qty	Total Cost
<u>6808374</u>	<u>GEARED ENGINE</u>	<u>1</u>	<u>N/C WARR</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP LABOR</u>		<u>1 HR</u>	<u>N/C WARR</u>

Repair Notes:  
ERROR 32 - MOTOR FAULTY - REPAIRED / FC IS WJ SERIES -  
REPLACE AS PM - CAL W/ QC & OPS CHECKS

Service Technician: [Signature]

Date: 10-04-08