

# Return & Repair Form:

## Customer Information:

Company Name: Tinton Falls PD B: \_\_\_\_\_ S: \_\_\_\_\_

Date Received: 4/30/2010 Date Given to Service: \_\_\_\_\_

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR XB - 0072  
 9510- Drug Tester 5000

Description: A - B - Plus - Screener - Demo Printer Ser #: AR -

Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_

Other \_\_\_\_\_ Probe Ser # DD P -

## ACCESSORIES:

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_

Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_

Mouthpieces  Carrying Case  Dry Gas

Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return:

FC

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C W</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

Repair Notes:

FC# ARAN-0862

Cal w/OL i Ops check

Service Technician: JS

Date: 04/30/2010