

# Return & Repair Form:

## Customer Information:

Company Name: Patterson P.D. B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 9-8-2010 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR X13-0073  
9510-Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: FC

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W</u>
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C W</u>
<u>6809458</u>	<u>IR - Detector</u>	<u>1</u>	<u>N/C W</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

Notes:

NEW FC # AR 13E-1087, Tech found a few spots on cassette, they were removed.

Cal w/ QC & ops check

Service Technician: [Signature]

Date: 9-9-2010