

# Return & Repair Form:

## Customer Information:

Company Name: Buena Vista P.D B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 8-17-2010 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: ~~7410-7110-6310-6810-7510~~ 9510-Drug Tester 5000 Serial #: AR 413-0024  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR - \_\_\_\_\_  
 Whole Instrument \_\_\_\_\_ Top 1/2 \_\_\_\_\_ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD \_\_\_\_\_ P - \_\_\_\_\_

**ACCESSORIES:**

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return:			
<u>CUVETTE, FC</u>			
Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W</u>
<u>6809509</u>	<u>CUVETTE</u>	<u>1</u>	<u>N/C W</u>
<u>6809512</u>	<u>3MM B DRASS</u>	<u>1</u>	<u>N/C W</u>
<u>MP CAL 21</u>	<u>CAL</u>	<u>1</u>	<u>N/C W</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

### Repair Notes:

FC # ARBD-2517

CUVETTE damage due to over heated simulator.  
Cal w/DC; DPS check

FC out of spec date, replaced FC.

Service Technician: [Signature]

Date: 8-19-2010