



# RETURN AND REPAIR FORM

## Customer Information

B:

S:

Company Name: Spring Lake P.D., Spring Lake, NJDate Received: 07/07/14Date given to service: 07/07/14Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STDProduct:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 WholeSerial Number: AR XB-0076Printer Serial#: ARSim Serial#: DDProbe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

 110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.Fuel Cell = ARFD-1950CAL W/QC AND OPS CHECKService Technician BCDate: 07/15/14