

# Return & Repair Form:

## Customer Information:

Company Name: Mahwah Pol. Dept. B: LO8946 S: LO8947  
 Date Received: 10.12.10 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-7110-6510-6810-7510  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo  
 Serial #: AR XB - 0078  
 Printer Ser #: AR  
 Whole Instrument \_\_\_\_\_ Top 1/2 \_\_\_\_\_  
 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date Jan 2010

## Repair Information:

Test #

Reason for Return:

Part Number	Description	Qty	Total Cost
<u>MP cal 71</u>	<u>Cal</u>	<u>1</u>	
<u>MP Labor</u>	<u>Labor</u>	<u>0.5</u>	

Repair Notes:

This unit was inspected by Joe Eckard and cleared for Cal.  
NO PROBLEMS WITH UNIT FOUND - SOME SMALL COSMETIC DAMAGE ON THE EXTERIOR OF CASE AND TOP BLUE COVER - 10-14-10  
New cap tube

cal w/DC; OPS check

Service Technician: \_\_\_\_\_

Date: 10/13/2010