

Return & Repair Form:

Customer Information:

Company Name: Sullivan P.D. B: _____ S: _____
 Date Received: 1-18-11 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR XB - 0085
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:			
Part Number	Description	Qty	Total Cost
<u>8315076</u>	<u>PCB</u>	<u>1</u>	<u>N/CW</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/CW</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>N/CW</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/CW</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/CW</u>

Repair Notes:

FC# ARBL-1715
CAL W/FC & OPS Check
New Captube
New PCB
Memory Near Full, replace PCB.

Service Technician: [Signature]

Date: 1/28/2011