

Return & Repair Form:

Customer Information:

Company Name: Englewood Cliffs Police B: _____ S: _____
 Date Received: 2-9-10 Date Given to Service: _____
 Carrier: RedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other _____

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR XB - 0086
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P - _____
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FL

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FL</u>	<u>1</u>	<u>Warranty</u>
<u>6809458</u>	<u>IR Detector</u>	<u>1</u>	<u>Warranty</u>
<u>MP cal 71</u>	<u>cal</u>	<u>1</u>	<u>Warranty</u>
<u>8312820</u>	<u>PUMP</u>	<u>1</u>	<u>Warranty</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>Warranty</u>

Repair Notes:

New FC# ARQM-0778 and IR Detector
cal w/DC & OPS check
Replaced old pump made void noise

Service Technician: JS

Date: 02/11/2010