

# RETURN AND REPAIR FORM



## Customer Information

B: \_\_\_\_\_ S: \_\_\_\_\_

Company Name: Tenafly Boro P.D., Tenafly, NJ

Date Received: 07/15/2016

Date given to service: 07/15/2016

Carrier:  FedEx  UPS  USPS      Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR XC-0005

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: December 2016

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC
4414166	Labor	2	NC

Repair Notes: Adjusted IRV which was too low and put it back within range limits. Recalibrated unit as well.  
Ran two linearity tests, one before adjustments and recalibration, and after adjustments and recalibration was done.

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 07/28/2016