

Return & Repair Form:

Customer Information:

Company Name: River Valley B: _____ S: _____
 Date Received: 3-22-11 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR XC -0065
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: _____

Part Number	Description	Qty	Total Cost
<u>8315075</u>	<u>PCB</u>	<u>1</u>	<u>w/cw</u>
<u>mp cal 71</u>	<u>Cal</u>	<u>1</u>	<u>w/cw</u>
<u>mp labor</u>	<u>Labor</u>	<u>.5</u>	<u>w/cw</u>

Repair Notes:

Unit came in because of corrupt memory.
Installed new PCB.
Cal w/ QC & ops check.

Service Technician: AS

Date: 3/23/2011