# RETURN AND REPAIR FORM

Dräger

## Customer Information

B: S:

**Company Name:** Red Bank Boro. P.D., Red Bank, NJ

**Date Received:** 12/21/2017 **Date given to service:** 12/21/2017

**Carrier:** ☐ FedEx ☐ UPS ☐ USPS **Shipping Method:** ☐ GRD ☐ 3DAY ☐ 2DAY ☐ NDA-PRI ☐ NDA-STD

**Product:** ☐ 6510 ☐ 6810 ☒ 7110 ☐ 8610 ☐ 7510 ☐ DT5000 ☐ 7410 Upper-half ☐ 9510 ☐ 7410 Whole

Serial Number: AR XC-0069

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires:

**Description:** ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

## Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify)

## Repair Information: Test#:

| Part Number | Description | Qty | Total Cost |
|---|---|---|---|
| 4414161 | Calibration | 1 | |
| 4414166 | Labor | .5 | |
| 6808455 | Fuel Cell | 1 | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

Repair Notes: Replaced defective fuel cell.

Fuel Cell = ARKA-1706

CAL W/QC AND OPS CHECK

**Service Technician** BC **Date:** 01/19/2018

sustxfs010\Public\Forms\return and repair form 07_21_2011.pdf