RETURN AND REPAIR FORM

Dräger

Customer Information

B: S:

Company Name: Mansfield Twp P.D., Columbus N.J.

Date Received: 02/28/12 Date given to service: 02/28/12

Carrier: [x] FedEx [ ] UPS [ ] USPS

Shipping Method: [ ] GRD [x] 3DAY [ ] 2DAY [ ] NDA-PRI [ ] NDA-STD

Product: [ ] 6510 [ ] 6810 [x] 7110 [ ] 8610 [ ] 7510 [ ] DT5000 [ ] 7410 Upper-half [ ] 9510 [ ] 7410 Whole

Serial Number: AR XC-0071

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires:

Description: [ ] A [ ] B [ ] Plus [ ] Demo [ ] Screener [ ] Trade In

Accessories

[ ] 110V A/C Adapter [ ] Regulator [ ] Printer Ribbon [ ] Printer Paper
[ ] Mouthpieces [ ] 9510 Stylus [ ] 9510 Top Cover [ ] Carrying Case
[ ] Dry Gas [ ] Other (please specify)

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
6808455	Fuel Cell	1	N/C W
MP CAL 71	Calibration	1	N/C W
MP Labor	Labor	.5	N/C W

Repair Notes: Replaced capacitor at location C167, to restore modem connection. Also installed a new Fuel Cell # MRCK-092310 replace busted Fuel cell # ARYM-2286.
Cal w/ QC & ops check

Service Technician: [signature] Date: 3/8/12

sustcfs010\Public\Forms\return and repair form 07_21_2011.pdf