RETURN AND REPAIR FORM

Dräger

Customer Information

B: S:

Company Name: Mansfield P.D., Columbus, NJ

Date Received: 12/16/2016 **Date given to service:** 12/16/2016

Carrier: ☐ FedEx ☒ UPS ☐ USPS **Shipping Method:** ☒ GRD ☐ 3DAY ☐ 2DAY ☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810 ☒ 7110 ☐ 8610 ☐ 7510 ☐ DT5000 ☐ 7410 Upper-half ☐ 9510 ☐ 7410 Whole

Serial Number: AR XC-0071

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires:

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify)

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	.5	
6808455	Fuel Cell	1	
6808486	Spacer Plates	2	

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARJE-1722

CAL W/QC AND OPS CHECK

Service Technician BC **Date:** 12/30/2016

sustxfs010\Public\Forms\return and repair form 07_21_2011.pdf