

Return & Repair Form:

Customer Information:

Company Name: Harrison P. D. B: _____ S: _____
 Date Received: 1-27-09 Date Given to Service: _____
 Carrier: FedEx UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 7110 - 6510 - 6810 Serial #: AR XC-0083
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: IR RESULTS

Part Number	Description	Qty	Total Cost
<u>6807458</u>	<u>IR DETECTOR PCB</u>	<u>1</u>	<u>N/C WTRAK</u>
<u>MPCAL F1</u>	<u>CAL</u>	<u>1</u>	<u>N/C WTRAK</u>
<u>IMPLABOR</u>		<u>.5</u>	<u>N/C WTRAK</u>

Repair Notes:

FILTER BUMP TEST OK - REPLACE DETECTOR PCB / FC IS 2D -
REPLACED 7-1-08
CAL w/ QC EOPS CHECKS

Service Technician: [Signature]

Date: 01-28-09 JS
01-29-09 JC