

# Return & Repair Form:

**Customer Information:** Eaton town P.P.

Company Name: IR Communications B: \_\_\_\_\_ S: \_\_\_\_\_

Date Received: 12-29-09 Date Given to Service: \_\_\_\_\_

Carrier: ~~FedEx~~ UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

**Product Information:**

Product: 7410-7110-6510-6810-7510 Serial #: AR XC - 0085

9510- Drug Tester 5000

Description: A - B - Plus - Screener - Demo Printer Ser #: AR -

Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_

Other \_\_\_\_\_ Probe Ser # DD P -

**ACCESSORIES:**

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_

Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_

Mouthpieces  Carrying Case  Dry Gas

Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

**Repair Information:**

Test #

Reason for Return: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
<u>6809514</u>	<u>Breath Hose</u>	<u>1</u>	<u>409.50</u>
<u>MPCAC</u>		<u>1</u>	<u>N/C W/MOR</u>
<u>MPLABOR</u>		<u>.5</u>	<u>44.00</u>

**Repair Notes:**

Unit arrived w/leak coming in breath hose, check cover the front screen no leak found, Replaced hose need to be changed. FL still in date range April, 2008. (installed 7/08)

Get w/DC's OPG check

Service Technician: AS

Date: 12/30/09