

RETURN AND REPAIR FORM



Customer Information

Company Name: Wyckoff, P.D. Wyckoff W.I.

Date Received: 12/06/11 Date given to service: \_\_\_\_\_

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  7110  8610  7510  DT5000  7410 Upper-half  9510  7410 Whole

Serial Number: AR XC-0090  
 Printer Serial#: AR  
 Sim Serial#: DD  
 Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	<u>N/C/W</u>
<u>mp Cal 71</u>	<u>Calibration</u>	<u>1</u>	<u>N/C/W</u>
<u>mp Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C/W</u>

Repair Notes: Replaced Fuel Cell ARYE-0563 with Fuel Cell ARCE-1315, because leakage.

Cal w/DC : ops check

Service Technician: JF Date: 12-8-11