RETURN AND REPAIR FORM

Dräger

Customer Information

B: S:

Company Name: Palisades Park Police Dept., Palisades Park, NJ

Date Received: 01/27/14 **Date given to service:** 01/27/14

Carrier: [X] FedEx [ ] UPS [ ] USPS **Shipping Method:** [ ] GRD [X] 3DAY [ ] 2DAY [ ] NDA-PRI [ ] NDA-STD

Product: [ ] 6510 [ ] 6810 [X] 7110 [ ] 8610 [ ] 7510 [ ] DT5000 [ ] 7410 Upper-half [ ] 9510 [ ] 7410 Whole

Serial Number: AR XC-0096

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: Jan. 2014

Description: [ ] A [ ] B [ ] Plus [ ] Demo [ ] Screener [ ] Trade In

Accessories

[ ] 110V A/C Adapter [ ] Regulator [ ] Printer Ribbon [ ] Printer Paper
[ ] Mouthpieces [ ] 9510 Stylus [ ] 9510 Top Cover [ ] Carrying Case
[ ] Dry Gas [ ] Other (please specify) _____

Repair Information

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
MP Labor	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W
6808374	Geared Engine (Motor)	1	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = AREL-1273

CAL W/QC AND OPS CHECK

Service Technician BC Date: 01/30/14

sustxf6010\Public\Forms\return and repair form 07_21_2011.pdf