

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Midland Park PD, Midland Park, NJ

Date Received: 11/21/11 Date given to service: _____

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR XC-0099
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information

Test#:

Part Number	Description	Qty	Total Cost
<u>MP Cal 71</u>	<u>Calibrations</u>	<u>1</u>	<u>N/C W</u>
<u>13336</u>	<u>Hose</u>	<u>9 IN</u>	<u>N/C W</u>
<u>mp labor</u>	<u>labor</u>	<u>.5</u>	<u>N/C W</u>

Repair Notes: The hose that connects the Pump and the solenoid had a crimp in the line, replaced hose.

Cal w/DC; ops check

Service Technician: [Signature]

Date: 11-21-11