

## RETURN AND REPAIR FORM



### Customer Information

B:

S:

Company Name: Closter P.D., Closter, NJ

Date Received: 12/07/2016 Date given to service: 12/07/2016

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  7110  8610  7510  DT5000  7410 Upper-half  9510  7410 Whole

Serial Number: AR XC-0102  
 Printer Serial#: AR  
 Sim Serial#: DD  
 Probe Serial#: DD

Warranty Expires: December 2016

Description:  A  B  Plus  Demo  Screener  Trade In

### Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

### Repair Information:

### Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	1	NC/W
6808455	Fuel Cell	1	NC/W

Repair Notes: Replaced defective fuel cell, c156, and c167.  
C156 and C167 were changed to resolve the modem issue.

Fuel Cell = ARJE-1822  
CAL W/QC AND OPS CHECK

Service Technician BC Date: 12/09/2016