

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: West New York P.D., West New York, NJ

Date Received: 01/14/14 Date given to service: 01/14/14

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR XC-0105  
 Printer Serial#: AR  
 Sim Serial#: DD  
 Probe Serial#: DD

Warranty Expires: Dec. 2013 (77868)

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	
MP Labor	Labor	1	

Repair Notes: Replaced defective C156 and C167 due to modem issues.

\_\_\_\_\_

\_\_\_\_\_

CAL W/QC AND OPS CHECK

\_\_\_\_\_

\_\_\_\_\_

Service Technician BC Date: 01/17/14