

Return & Repair Form: 10382112

80488734

Customer Information:

Company Name: Glenn Rock P.D. B: 77933 S:

Date Received: 12-19-09 Date Given to Service:

Carrier: FedEx UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR XD - 0001
9510- Drug Tester 5000

Description: A - B - Plus - Screener - Demo Printer Ser #: AR

Whole Instrument Top 1/2 Sim Ser #:

Other Probe Ser # DD P

ACCESSORIES:

- 110 V A/C Adapter
- Regulator
- Mag Card Rdr#
- Printer Paper
- Printer Ribbon
- Casio #
- Mouthpieces
- Carrying Case
- Dry Gas

Other (Please Specify) Warranty Exp. Date

Repair Information:

Reason for Return: Test # FL

| Part Number | Description | Qty | Total Cost |
|-------------|-------------|-----|------------|
| 6808453 | FL | 1 | w/c war |
| MP CAL 71 | CAL | 1 | w/c war |
| MP Labor | Labor | 5 | w/c war |

Repair Notes: Install NEW FC# ARAM-0195

CAL w/DC; Ops check

Service Technician: JS

Date: 12/26/09