

Return & Repair Form:

Customer Information:

Company Name: Allendale Police B: _____ S: _____
 Date Received: 11-17-09 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR XD - 0002
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Reason for Return: ERROR 9 - PROBLEM IS ON MAIN PCB Test # _____

Part Number	Description	Qty	Total Cost
<u>8315075</u>	<u>PCB</u>	<u>1</u>	<u>N/C/WARR</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C/WARR</u>
<u>MP Labor</u>	<u>Labor</u>	<u>1hr</u>	<u>N/C/WARR</u>

Repair Notes:

DATA DOWN LOAD VIA MODEM TO TPR ROBERT DEMIER
11-18-09
Installed New PCB
Cal w/QC, ops check

Service Technician: JS

Date: 12/01/09