

RETURN AND REPAIR FORM



Customer Information

B: _____ S: _____

Company Name: New Milford P.D., New Milford, NJ

Date Received: 02/01/2017 Date given to service: 02/01/2017

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810 7110 8610
 7510 DT5000 7410 Upper-half 9510 7410 Whole

Serial Number: AR XD-0005
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#: _____

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	.5	
6808455	Fuel Cell	1	
6808486	Spacer Plates	2	

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARJL-1016
CAL W/QC AND OPS CHECK

Service Technician BC Date: 03/01/17