

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Guttenberg P.D., Guttenberg NJ.

Date Received: 04/03/12 Date given to service: 04/03/12

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  7110  8610  7510  DT5000  7410 Upper-half  9510  7410 Whole

Serial Number: AR KD-0006  
 Printer Serial#: AR  
 Sim Serial#: DD  
 Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	<u>N/C W</u>
<u>mp cal 71</u>	<u>Calibration</u>	<u>1</u>	<u>N/C W</u>
<u>mp Labor</u>	<u>Labor</u>	<u>5</u>	<u>N/C W</u>

Repair Notes: Replaced busted Fuel cell # AR2D-2424JS  
new Fuel cell # ARcm-2367. 2441

cal w/DC; ops check

Service Technician: [Signature] Date: 4-5-12